



Emotion® pink nitrile gloves

The instruction below should be used in conjunction with detailed information on the packaging.

Short description of the product	Properties / Application
examination and protective gloves, nitrile, powder-free, non-sterile for disposable use	These are non-sterile examination and protective gloves for single use, intended for use in medical field to: protect patient and user from cross-contamination, conducting medical examinations, diagnostic and therapeutic procedures and for handling medical contaminated material. Gloves are classified as Medical Devices Class I and as a Personal Protective Equipment category I simple design – for minimal risk. Their design and labelling corresponds to the requirements of the European Medical Device Directive 93/42/EEC and the European Directive on Personal Protective Equipment 89/686/EEC, both as amended. It should be noted, that confirmed protection level does not reflect real working conditions, where other factors may influence to the protection level. Gloves should be used solely according to their intended application.
Full description of the product	
raw material: nitrile external surface: fingertips textured internal surface: polymerized + chlorinated cuff: beaded colour: pink shape: ambidextrous, fitting to the right and left hand size range XS (5-6), S (6-7), M (7-8), L (8-9), XL (9-10) AQL: 1.5 Quantity in an individual packaging: 100 pcs. by weight Shelf life: 3 years (from the date of manufacturing) Minimum glove length – 240mm – gloves suitable for special purposes, there are examination gloves.	
Storage instructions	Contact with food
It is recommended to store the gloves in a dry place, in the temperature of 5-35°C and to protect them against sunlight and fluorescent light. Recommended relative humidity in the room where the gloves are stored is 60 ±20%. Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone. Do not keep in direct vicinity of solvents, oils, fuels and lubricants.	 Gloves marked with the symbol indicating contact with food comply with the requirements of Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Not all gloves can be used in contact with every type of food, because some of them can release harmful substances penetrating directly into the food. For more detailed information concerning intended use to particular food category, please contact the company Mercator Medical.
Precautions and indications for use	Components / hazardous components
Dry the hands well before putting the gloves on. Check the gloves before wearing them if they do not have any faults (holes, contaminations). After taking the gloves off, wash the hands thoroughly, dry them and apply a nourishing cream. Use at least 1 pair of gloves for one patient and one procedure, these are disposable gloves. The gloves soiled with blood or body fluids should be disposed as medical waste. Take care not to let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects. Protection is limited only to palm, results relate to palm and were examined in laboratory conditions. With regard to electrostatic properties it is pointed out that all garments and shoes wearied with gloves should be designed concerning electrostatic danger.	Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. Gloves made from natural rubber latex may cause allergic reactions, including anaphylactic responses. In the case of an allergic reaction, seek medical assistance immediately.
	Disposal
	Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.
	Manufacturer
	EFALOCK Professional Tools GmbH 97092 Würzburg Germany www.efalock.de
CE classification - Product Standards Compliance	
Gloves are classified as Medical Device – class I according to Council Directive 93/42/EEC. Compliance with standards: EN 455-1:2000, EN 455-2:2009+A2:2013, EN 455-3:2006, EN 455-4:2009, EN 980:2008, EN 1041:2008.	
Gloves are classified as Personal Protective Equipment – category I for minimal risk according to Council Directive 89/686/EEC. Compliance with standards: EN 420:2003+A1:2009,	